

## **CERTIFICATE OF ANALYSIS**

Product Code	Name		Quantity	
SI 390.900	URO-QUICK SCREEN	manufacture 27	d	
				 -
Lot Number	Manufacturing Da	ate Expiry Date	CAP color	Printed Band
SCR23270	26/07/2023	3 23/09/2024	Green	White

## **Eugonic Broth**

Lot Number SCR23270		Physical parameters (Volume, transparency, colour )	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expiry Date 23/09/2024	Expected	Compliant	No growth	CV%<=35%	CV%<=35%
CAP color Green Printed Band White	Found	ОК	ОК	29.9%	< 5%

Check of the presenze of the magnetic stirrer in each vial contained in the kit: YES

Storage Conditions: min 4°C - Max 30°C

Compliance: YES

Controlled By	Cirio Alberto	Date of Control	31/07/2023
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	31/07/2023

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance ...... Falcone Francesco

Date of Release 31/07/2023

MIC COA \$1390900#\$CR23270 pdf

This is a computer-generated document. No signature is required.

## **Registered Office**

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com