



# CERTIFICATE OF ANALYSIS

MIC\_COA\_SI390900#SCR14189.pdf

|                                   |   |                                    |  |
|-----------------------------------|---|------------------------------------|--|
| Product Code<br><b>SI 390.900</b> | Name<br><b>URO-QUICK SCREENING KIT 360 TEST</b> | Quantity manufactured<br><b>26</b> |  |
|-----------------------------------|---|------------------------------------|--|

|                               |   |                                  |                           |                              |
|-------------------------------|---|----------------------------------|---------------------------|------------------------------|
| Lot Number<br><b>SCR14189</b> | Manufacturing Date<br><b>04/08/2014</b> | Expiry Date<br><b>04/10/2015</b> | CAP color<br><b>Green</b> | Printed Band<br><b>White</b> |
|-------------------------------|---|----------------------------------|---------------------------|------------------------------|

## Eugonic Broth

Lot Number **SCR14189**  
 Expiry Date **04/10/2015**  
 CAP color **Green**  
 Printed Band **White**

|          | Physical parameters<br><i>(Volume, transparency, colour)</i> | Decontamination at<br>37°C after 72 hours | Growth inter batch | Growth intra batch |
|----------|--|---|--------------------|--------------------|
| Expected | Compliant  | No growth                                 | CV% <= 35%         | CV% <= 35%         |
| Found    | OK   | OK  | 12.6%              | < 5%               |

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Date of Control **26/08/2014**

Date of Approvation **26/08/2014**

Tested **Cirio Alberto**

Quality Control Manager **Cossettini Marco**

**ALIFAX**  
 via Petrarca 2/1 - 35020 POLVERARA (PD) - ITALY  
 Tel. +39 0490992000  
 Web site: [www.alifax.com](http://www.alifax.com) e-mail: [info@alifax.com](mailto:info@alifax.com)



**SIRE Analytical Systems S.r.l.**  
 via Merano 30 - 33045 Nimis (UDINE) - ITALY  
 Tel. +39 0432547454