



CERTIFICATE OF ANALYSIS

MIC_COA_SI390900#SCR14278.pdf

Product Code SI 390.900	Name URO-QUICK SCREENING KIT 360 TEST	Quantity manufactured 26	
-----------------------------------	---	------------------------------------	--

Lot Number SCR14278	Manufacturing Date 28/11/2014	Expiry Date 03/02/2016	CAP color Green	Printed Band White
-------------------------------	---	----------------------------------	---------------------------	------------------------------

Eugonic Broth

Lot Number **SCR14278**
 Expiry Date **03/02/2016**
 CAP color **Green**
 Printed Band **White**

	Physical parameters <i>(Volume, transparency, colour)</i>	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% <= 35%	CV% <= 35%
Found	OK	OK	11.9%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Date of Control **10/12/2014**

Date of Approval **10/12/2014**

Tested **Cirio Alberto**

Quality Control Manager **Cossettini Marco**

ALIFAX
 via Petrarca 2/1 - 35020 POLVERARA (PD) - ITALY
 Tel. +39 0490992000
 Web site: www.alifax.com e-mail: info@alifax.com



SIRE Analytical Systems S.r.l.
 via Merano 30 - 33045 Nimis (UDINE) - ITALY
 Tel. +39 0432547454