

CERTIFICATE OF ANALYSIS

MIC COA SI200000#SCD20115 pdf

				90900#SCR20115.pdf
Product Code	Name	Quantity manufactured		
SI 390.900	KIT URO QUICK SCREENING 36	27		
Lot Number	Manufacturing Date	Expiry Date	CAP color	Printed Band
SCR20115	15/05/2020	14/07/2021	Green	White

Eugonic Broth

Lot Number	SCR20115		Physical parameters (Volume, transparency, colour)	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expiry Date	14/07/2021	Expected	Compliant	No growth	CV%<=35%	CV%<=35%
CAP color	Green	Found	ОК	ОК	27.5%	< 5%
Printed Band	White					

Check of the presenze of the magnetic stirrer in each vial contained in the kit: YES

Storage Conditions: min 4°C - Max 30°C

Compliance:

YES

Controlled By	Cirio Alberto	Date of Control	22/05/2020
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	22/05/2020

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

 Released By Quality
 Zorzut Milena
 Date of Release
 22/05/2020

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com