

CERTIFICATE OF ANALYSIS

MIC_COA_SI390900#SCR20108.pdf Quantity Product Code Name manufactured SI 390.900 **KIT URO QUICK SCREENING 360 TEST** 27 Manufacturing Date Lot Number Expiry Date Printed Band CAP color SCR20108 06/07/2021 Green White 07/05/2020

Eugonic Broth

Lot Number	SCR20108		Physical parameters (Volume, transparency, colour)	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expiry Date	06/07/2021	Expected	Compliant	No growth	CV%<=35%	CV%<=35%
CAP color	Green	Found	ОК	ОК	14.8%	< 5%
Printed Band	White					

Check of the presenze of the magnetic stirrer in each vial contained in the kit: YES

Storage Conditions: min 4°C - Max 30°C

Compliance:

YES

Controlled By	Cirio Alberto	Date of Control	14/05/2020
Approved By Manufacturing Manager	Pividori Denis	Date of Approval	15/05/2020

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

Released By Quality	Falcone Francesco	Date of Release	15/05/2020

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com