

## **CERTIFICATE OF ANALYSIS**

MIC\_COA\_SI390900#SCR22074.pdf

Product Code	Name	Quantity manufactured	
SI 390.900	URO-QUICK SCREENING KIT	27	

Lot Number	Manufacturing Date	Expiry Date	CAP color	Printed Band
SCR22074	22/03/2022	21/05/2023	Green	White

## **Eugonic Broth**

Lot Number SCR22074
Expiry Date 21/05/2023

CAP color **Green**Printed Band **White** 

	Physical parameters (Volume, transparency, colour )	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%<=35%	CV%<=35%
Found	OK	OK	19.7%	7.4%

Check of the presenze of the magnetic stirrer in each vial contained in the kit: YES

Storage Conditions: min 4°C - Max 30°C

Compliance: YES

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. These requirements have been established according to the European Directive 98/79/EC. We confirm that the above referenced product conformed to the specifications and then it is released.

This is a computer-generated document. No signature is required.

**Registered Office** 

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



**Manufacturing Site** 

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com