

Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321 13088 Berlin, Germany

T: +49 30 927 030-0 F: +49 30 927 030-30 info@sifin.de www.sifin.de

Product

Anti-Coli O 86:K 61

Catalogue no.

TS 2206

Package size

1 ml

Lot no.

18 01 16

Expiry date

2021-01

Intended use

Anti-Coli O 86:K 61 records all E. coli strains that contain the antigen

O 86:K 61 in slide agglutination. In the Widal reaction the serum reaches a titre of 1:800 (tube and MTP) with the homologous E. coli strains; deviations of ± 1

titre step are acceptable.

Specificity

Since unspecific co-agglutinations with other E. coli strains often occur in the

slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled *E. coli* isolate in the Widal test.

Preservatives

Sodium azide 0.9 mg/ ml

Storage

2 °C...8 °C

Berlin. 09.03.46

place, date

signature Head of Quality Control