

## Certificate of Quality Control

CERTIFIED QUALITY MANAGEMENT SYSTEM IN  
 ACCORDANCE WITH DIN EN ISO 9001 AND DIN EN ISO 13485

sifin diagnostics gmbh

Berliner Allee 317-321  
 13088 Berlin, Germany

T: +49 30 927 030-0  
 F: +49 30 927 030-30  
 info@sifin.de  
 www.sifin.de

|                    |   |
|--------------------|---|
| <b>Product</b>     | <b>Anti-Coli O 118:K-</b>   |
| Catalogue no.      | TS 2220   |
| Package size       | 1 ml  |
| <b>Lot no.</b>     | <b>179 08 15</b>  |
| <b>Expiry date</b> | <b>2020-08</b>  |
| Intended use       | Anti-Coli O 118:K- records all <i>E. coli</i> strains that contain the antigen O 118:K- in slide agglutination. In the Widal reaction the serum reaches a titre of 1:3200 (tube and MTP) with the homologous <i>E. coli</i> strains; deviations of $\pm 1$ titre step are acceptable. |
| Specificity        | Since unspecific co-agglutinations with other <i>E. coli</i> strains often occur in the slide agglutination test, a positive result in the slide test must always be confirmed by examination of the boiled <i>E. coli</i> isolate in the Widal test.                                 |
| Preservatives      | Sodium azide 0.9 mg/ ml   |
| Storage            | 2 °C.. 8 °C   |

Berlin, 09.10.15  
 place, date

  
 signature Head of Quality Control