



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: **Xpert® Xpress Flu/RSV**

Cepheid Catalogue Part No.:

Kit Lot No.:

Cartridge Lot No.:

Kit Expiration Date:

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna Sunnyvale

Functional Testing

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Negative	Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE	Passed
High Positive	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE	Passed

20190114

Signature of Quality Assurance

Date

Name:

Title: