

High Positive

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name:	Xpert® Xpress Flu/RSV			
Cepheid Catalogue P	art No.: XPRSFLU/RSV-CE-10]		
Kit Lot No.: 1000	173887			
Cartridge Lot No.: 11	907			
Kit Expiration Date:	2021-03-07			
Legal Manufacturer		Manufacturing Facility	Solna	← Sunnyvale
Cepheid		Cepheid AB	,	Camiyvaic
904 Caribbean Drive	10.4	Röntgenvägen 5		
Sunnyvale, CA 94089 U	JSA	SE-171 54 Solna Sweden		
		Sweden		
Functional Testing				
Test Description	Acceptance Criteria		Test Result	
Negative	Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE		Passed	
Low Positive	Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE		Passed	

Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE

Signature of Quality Assurance

Name: Samaneh Vahid

Title: QA Analyst

Passed