

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
Doxycycline DXT 0.016-256 µg/mL	111921001	19.11.2021	2024.11.18

**Ref.**

92156 – 921560 – 921561

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–0.5	0.12
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	2–8	6
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	1
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.015–0.12	0.064
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	2–8	3
<i>Eggerthelia lenta</i> ATCC® 43055	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	2–16	8
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO <sub>2</sub> 16-20 h	0.25–1**	0.38

\*Mueller Hinton II (Cation Adjusted) Agar: Liofilchem, ref. 10031; BD-BBL, ref. 212257

\*\*Established and validated by EUCAST

**Batch Release**

Approved

Date

30.11.2021

Signature

Quality Control

(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.