

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® HCV Viral Load				
Cepheid Catalogue Part No.: GXHCV-VL-CE-10				
Kit Lot No.: 1000325884				
Cartridge Lot No.: 35803				
Kit Expiration Date: 2022-12-11				
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Solna ○ S	Sunnyvale
Functional Testing according to D14282 Rev.: M				
Test Description	Acceptance Criteria		Test Result	
LOW Positive	HCV DETECTED: 1.50-1.90 log ₁₀ IU/mL		Passed	
HIGH Positive	HCV DETECTED: 4.25-4.55 log ₁₀ IU/mL			Passed
Negative	HCV NOT DETECTED			Passed
Signature of Quality Assurance Date				
Name: Robert Fiedler				
Title: QA Analyst				