

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Amphotericin B AMB 20 µg	120921103	2024.12.08

Ref.

9071 – 9071/1 – 9071/2

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	18
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	19
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	17
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	18

Batch Release

Approved

Date

17.12.2021

SignatureQuality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.