

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Metronidazole MTZ 5 µg	012722052	2025.01.26

Ref.

9076 – 9076/1 – 9076/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Bacteroides fragilis</i> ATCC 25285	Schalder K Agar (Sheep Blood 5%)	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h		38
<i>Bacteroides thetaiotaomicron</i> ATCC 29741	Schalder K Agar (Sheep Blood 5%)	0.5 McFarland	35 ± 1°C, 5% CO ₂ 16-20 h		35

Batch Release

Approved

Date

01.02.2022

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.