

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Fosfomycin FOS 200 µg	032522042	2025.03.24

**Ref.**

9109 – 9109/1 – 9109/2

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	22-30	30
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	26-34*	30
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	25–33	30

\*Established and validated by EUCAST

**Batch Release**

Approved

**Date**

31.03.2022

**Signature**

Quality Control

(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.