

# Liofilchem®

## Certificate of Analysis

Product	Batch	Expiration date
Fluconazole FLU 25 µg	011822135	2025.01.17

**Ref.**

9166 – 9166/1 – 9166/2

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	28-39 28
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	22-33 26
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	-- 11
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylene Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	26-37 26

**Batch Release**

Approved

**Date**

25.01.2022

**Signature**Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.