

CERTIFICATE OF ANALYSIS

| | | |
|--|------------------------------------|--|
| Product Name: Streptococcal Grouping Reagent Latex Suspension Group G | Lot Number: C19778F | |
| Product Code: PL.036 | Expiration Date: 2023-03-31 | |

Additional Information


| | |
|---------------------|-------------------------------|
| Storage Conditions: | 2-8°C |
| Precautions: | Refer to Instructions for Use |

Quality Control Testing

| Appearance | Reaction | Result |
|---|----------|--------------|
| Homogenous suspension with no agglutination | N/A | Satisfactory |
| Organisms | | |
| <i>Streptococcus</i> sp. ATCC 12394 (Group G) | Positive | PASS |
| <i>Streptococcus</i> Group G - Clinical Isolate | Positive | PASS |
| <i>Streptococcus pyogenes</i> ATCC 19615 (Group A) | Negative | PASS |
| <i>Streptococcus agalactiae</i> ATCC 12386 (Group B) | Negative | PASS |
| <i>Streptococcus equisimilis</i> ATCC 12388 (Group C) | Negative | PASS |
| <i>Enterococcus faecalis</i> ATCC 49477 (Group D) | * | PASS |
| <i>Streptococcus</i> sp. ATCC 12392 (Group F, type 2) | Negative | PASS |
| <i>Enterococcus faecalis</i> ATCC 19433 | Negative | PASS |
| <i>Enterococcus faecalis</i> ATCC 49478 | Negative | PASS |
| <i>Enterococcus faecalis</i> ATCC 29212 | Negative | PASS |
| <i>Listeria monocytogenes</i> NCTC 11994 | Negative | PASS |
| Control | | |
| Polyvalent Positive Control (PL.040) | Positive | PASS |

*Cross reaction may be present.

Release

| | |
|------------|--|
| 2021 10 22 |  |
| Date | Quality Control |

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com