



CERTIFICATE OF ANALYSIS

POSITIVE CONTROL KIT

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT
BATCH N.: PCK22004	EXPIRY DATE: 21/08/2023
PACKING DATE: 27/06/2022	PACKED KITS No.: 30

◆ RECONSTITUTING AND DILUTING SOLUTIONS

Eugonic broth

Batch N.: FLA22003

Expire date: 11/06/2024

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	OK	OK	15,9%	<5%

* Volume, transparency, colour

◆ LYOPHILIZED MICROORGANISM

Batch N.: 21002

Expire date: 19/10/2023

	Growth inter batch (CV% ≤35%)	Physical check**
Expected	Compliant	Compliant
Found	OK	OK

** Aspect

STORAGE CONDITIONS: 2 ÷ 8 °C

Date of Approval:
04/07/2022

Pividori Denis:
Manufacturing Manager

Date of Release:
04/07/2022

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance

ALIFAX S.r.l.

Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

©Copyright 2022. This document is exclusive property of ALIFAX S.r.l. confidential and not freely disclosable. All rights reserved.