



www.pro-lab.com

### CERTIFICATE OF ANALYSIS

<b>Product Name:</b> Proflow H.pylori	<b>Lot Number:</b> P-297
<b>Product Code:</b> PL.3006	<b>Expiration Date:</b> 2024 03 31

#### Additional Information

Test conditions: As defined in the IFU.	Storage conditions: 2-30°C	The product is intended for in vitro use only.
---	----------------------------	--

#### Quality Control Testing

Functionality test results: visual interpretation at 10 minutes

- 1. Sensitivity test.  
Dilutions from standard: approved.
- 2. Specificity test.  
Materials prepared in product buffer: approved.

#### Quality Control result

Approved.

#### Release

2022/03/31	<i>mawr .</i>
Date	Quality Control

Canada ☎ 800 268 2341 Fax 905 731 0206  
 20 Mural Street, Unit #4, Richmond Hill, Ontario,  
 L4B 1K3  
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424  
 21 Cypress Blvd., Suite 1155, Round Rock,  
 Texas 78665-1046  
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614  
 3 Bassendale Road, Bromborough, Wirral,  
 Merseyside CH62 3QL  
 uksupport@pro-lab.com