

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product  | Batch     | Production date | Expiration date |
|--|-----------|-----------------|-----------------|
| <b>Posaconazole POS</b><br><b>0.002-32 µg/mL</b> | 033022037 | 30.03.2022      | 2025.03.29      |

### Ref.

92152 – 921520 – 921521

### Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains                            | Medium        | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|--|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida krusei</i><br>ATCC® 6258        | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.125-0.5                            | 0.19                 |
| <i>Candida parapsilosis</i><br>ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.032-0.25                           | 0.064                |
| <i>Candida albicans</i><br>ATCC® 90028     | R.P.M.I. Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>24-48 h | 0.032-0.125                          | 0.064                |

### Batch Release

Approved

Date

29.04.2022

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.