

## QUALITY CONTROL CERTIFICATE

### FLUCONAZOLE 25 µg Neo-Sensitabs

REF: 82512

LOT: 220021

Date of analysis: 2022-09

Date of expiry: 2027-09

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Candida Paraposilosis* ATCC 22019  
*Candida albicans* ATCC 90028

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica a/s



Peter Samuelson  
Quality Assurance Manager