

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Vancomycin/Teicoplanin VA/TEC 0.5-32/0.5-32 µg/mL	051722065	17.05.2022	2025.05.16

Ref.

92163 – 921630 – 921631

Antimicrobial Susceptibility Testing

Mueller Hinton II Agar + 5% Sheep Blood, 0.5 McFarland in broth, ambient, 35 ± 2°C, 18-48 h

Control strains	Expected Results	Expected Results	Results
	Vancomycin MIC range (mg/L)	Teicoplanin MIC range (mg/L)	
<i>Staphylococcus aureus</i> ATCC® 29213 (negative control)	0.5–2	1–4	Conforms
<i>Staphylococcus aureus</i> ATCC® 700698 (h-GISA)	1–8	≥32	Conforms
<i>Staphylococcus aureus</i> ATCC® 700699 (GISA)	4–16	≥32	Conforms

Batch Release

Approved

Date

27.05.2022

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.