

Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Neomycin N 30 µg	111722047	2025.11.16

Ref.

9032 – 9032/1 – 9032/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	17-23 (CLSI) 14-20 (EUCAST)	20
<i>Staphylococcus aureus</i> ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	18-26	25
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16-22	25

Batch Release

Approved

Date

22.11.2022

SignatureQuality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.