

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Ampicillin AMP 0.016-256 µg/mL	012723001	27.01.2023	2026.01.26

Ref.

92003 – 920030 – 920031

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	0.5
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	0.75
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	2–8	6
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	>32	>32
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	>128	>128
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	2–8	3
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.06–0.25	0.094
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.06–0.25*	0.25

*Established and validated by EUCAST

Batch Release

Approved

Date 20.02.2023

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.