

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® CT/NG			
Cepheid Catalogue Part No.: GXCT/NG-CE-10			
Kit Lot No.: 1000170213			
Cartridge Lot No.: 31315			
Kit Expiration Date: 2021-02-28			
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden	Manufacturing Facility Solna Solna Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Sunnyvale	
Functional Testing  Test Description	Acceptance Criteria	Test Result	
Positive	CT detected; NG detected	Passed	
Negative	CT not detected; NG not detected	Passed	
Signature of Quality Assurance  Name: Marie Bah  Title: QA Manager Release			