

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
Teicoplanin TEC 0.016-256 µg/mL	050323002	03.05.2023	2026.05.02

**Ref.**

92012 – 920120 – 920121

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	1
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.75

**Batch Release**

Approved

**Date**

10.05.2023

**Signature**Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.