

• Certificate of Analysis

Product Name: **Bruker Matrix HCCA, portioned**

Part No.: 8255344

LOT No.: 6020122140

Use by¹⁾: 2024-09

Date of Production: 2022-10-26

Storage Conditions : **Store at 2°C to 8°C on arrival.**

| <u>Test:</u> | <u>Specification</u> | <u>LOT Results</u> |
|--|--|----------------------|
| 1. Amount of HCCA content. Sampling in accordance to DIN ISO 3951-1. | AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg | Passed |
| 2. MALDI Biotyper identification of a 1:20 diluted BTS. Sampling in accordance to DIN ISO 2859-1. | Classification Result Report lists " <i>Escherichia coli</i> " as best match. AQL-value=1.0 Inspection level S4 L= 2.000 log(score) of the best match | Passed Passed |
| 3. Blank value analysis of HCCA-portioned by MALDI Biotyper. Sampling in accordance to DIN ISO 2859-1. | Classification Result Report lists "no reliable identification". AQL-value=1.0 Inspection level S4 U= 1.699 log(score) | Passed Passed |

Remarks:

- 1) Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
28359 Bremen, Germany
phone +49-421-2205-0



BRUKER
Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
D-28359 Bremen
Dr. Thomas Maier, SAFETY OFFICER
Fax +49 (0)421 - 2205-370

QC release date: 28-10-2022

Not for use in diagnostic procedures. For research use only.