

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert <sup>®</sup> Carba-R						
Cepheid Catalogue Part No.: GXCARBARP-CE-10						
Kit Lot No.: 1000579115		3				
Cartridge Lot No.: 10501						
Kit Expiration Date: 2024-05-12						
Legal Manufacturer	Manufacturing Facility	Solna	○ Sunnyvale			
Cepheid 904 Caribbean Drive Supported CA 04080 USA	Cepheid AB Röntgenvägen 5					
Sunnyvale, CA 94089 USA	SE-171 54 Solna					

## Functional Testing

Test Description	Acceptance Criteria	Test Result
	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

Sweden

Anton Andersson	2022-11-20
Signature of Quality Assurance	Date
Name: Anton Andersson Engström	
Title: QA Analyst	