

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Polymyxin B PB 300 IU	030123112	2026.02.28

Ref.

9120 – 9120/1 – 9120/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	13–19	18
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	14–18	18

Batch Release

Approved

Date

07.03.2023

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.