

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® HCV VL Fingerstick				
Cepheid Catalogue Part No.: GXHCV-FS-CE-10				
Kit Lot No.: 1000180074				
Cartridge Lot No.: 12501				
Kit Expiration Date: 2020-09-20				
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Solna	Sunnyvale
Functional Testing according to D27172 Rev.: D				
Test Description	Acceptance Criteria			Test Result
LOW Positive	HCV DETECTED: 2.29-2.67 log ₁₀ IU/mL			Passed
HIGH Positive	HCV DETECTED: 5.50-5.90 log ₁₀ IU/mL			Passed
Negative	HCV NOT DETECTED			Passed
Signature of Quality Assurance Date Name: Marie Bah				
Title: QA Manager Release				