

## • Certificate of Analysis

Product Name : **Bruker Matrix HCCA-portioned**

Part No.: **8255344**

LOT No.: **6020123072**

Use by<sup>1)</sup>: **2025-05-31**

Date of Production: **2023-06-01**

Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
<b>1. Amount of HCCA content</b>	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
<b>2. MALDI Biotyper identification of a 1:20 diluted BTS</b>	Classification Result Report lists " <i>Escherichia coli</i> " as best match with a log score value $\geq 2.00$ .	Passed
<b>3. Blank value analysis of Matrix HCCA-portioned by MALDI Biotyper</b>	Classification Result Report lists "No Organism Identification Possible." with a log score value $< 1.70$ .	Passed

### Remarks:

1) Bruker Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG  
Fahrenheitstr. 4  
28359 Bremen  
Germany  
Phone +49-421-2205-0

QC release date: 2023-07-04

  
Quality Management