

## **CERTIFICATE OF ANALYSIS**

Product Name:	Prolex™ Staph Xtra Latex Kit L			Number:	C21787
Product Code:	PL.1080 Expiration Date:		on Date:	2025-02-28	
Product Code	Quantity	Compone	nts		
PL.1083	2 x 2.5 ml	Staph Xtra	a Latex Reagent		
PL.1085	1 x 5.0 ml	Staph Xtra	Staph Xtra Negative Control Latex Reagent		
	200	Mixing Stic	ks		
	20 cards	Test Cards			
Additional Information					
Storage Conditions: 2-8°C					
Instructions for Use: Refer to pro-lab.com/resources for current version					
Quality Control Testing					
Criteria					Result
			-		nooun
Strong agglutination of Staph Xtra Latex Reagent when tested against <i>Staphylococcus aureus</i> and MRSA* including capsular types 5 and 8 within 20 seconds.					PASS
5					
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.					PASS
* MRSA - (Methicil	lin-resistant Staph	vlococcus aureus	5)		
Release		-			
2023-07-10			Lara	lei	
Date			Quality Control		

U.S.A. <sup>@</sup> (512) 832 9145 Fax (512) 832 6424 21 Cypress Blvd., Suite 1155, Round Rock, Texas 78665-1046 support@pro-lab.us