

## **CERTIFICATE OF ANALYSIS**

ESR\_COA\_SI305#2439A.pdf

Product Code SI 305.300-A	LATI	Name EX CONTROLS	Quantity manufactured <b>178</b>	Type of Tubes  Greiner	
Lot Number		Manufacturing Date	Expiry Date		
2439A		04/10/2023	01/04/2024		

	Standar Transmittance	ESF	es R (mm in-Val-n			R (mm,		M Instrument	easured	Values	Serial Number	ОК		Volume checked
Level 2	1534	6	-	11	9	9	9	TEST1 THL			6608	X	NO	YES
Level 3	980	15	-	22	19	20	20	TEST1 THL			T13290	X	NO	YES
Level 4	637	56	-	<i>74</i>	65	67	68	ROLLER 20 PN			133	X	NO	YES

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance ...... Gaddi Giulia Date of Release 05/10/2023

This is a computer-generated document. No signature is required.

**Registered Office** 

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy Tel (+39) 049 0992000 - Fax (+39) 049 5855434 Mail PEC: info.alifax@pec.alifax.com info@alifax.com - www.alifax.com



**Manufacturing Site** 

Via Merano, 30 - 33045 Nimis (UD) Italy Tel (+39) 0432 547454 - Fax (+39) 0432 547378 Mail PEC: info.udine@pec.alifax.com info.udine@alifax.com