

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
<b>Cefotaxime CTX 0.002-32 µg/mL</b>	090423011	04.09.2023	2026.09.03

**Ref.**

**92007 – 920070 – 920071**

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	2
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.03–0.12	0.064
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32	12
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.12–0.5	0.25
<i>Haemophilus influenzae</i> ATCC® 49766	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.004–0.016	0.012
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.03–0.12	0.094
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO <sub>2</sub> 20-24 h	0.015–0.06	0.016

**Batch Release**

Approved

**Date** 11.09.2023

**Signature**

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.