

QUALITY CONTROL CERTIFICATE

AMPHOTERICIN B 10 µg Neo-Sensitabs

REF: 81012

LOT: 230020

Date of analysis: 2023-12

Date of expiry: 2026-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida albicans ATCC 90028
Candida parapsolis ATCC 22019
Candida krusei ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Manager Quality Assurance

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance