

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|--------------------------------|-----------|-----------------|-----------------|
| Fosfomycin FOS 0.016-256 µg/mL | 020124018 | 01.02.2024 | 2027.01.31 |

Ref.

92078 – 920780 – 920781

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---------------------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.5–4 | 0.5 |
| <i>Enterococcus faecalis</i> ATCC® 29212 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 32–128 | 96 |
| <i>Escherichia coli</i> ATCC® 25922 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.5–2 | 1 |
| <i>Pseudomonas aeruginosa</i> ATCC® 27853 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 2–8 | 2 |

Batch Release

Approved

Date

06.02.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.