

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name:	Xpert® HBV Viral Load				
Cepheid Catalo	gue Part No.: GXHBV-VL-CE-10	▼			
Kit Lot No.:	1001416415				

Kit Expiration Date: 2025-02-16

Cartridge Lot No.:

Legal Manufacturer	Manufacturing Facility	Solna Sunnyvale Newark
Cepheid AB	Cepheid AB	
Röntgenvägen 5 SE-17154 Solna	Röntgenvägen 5	
Sweden	SE-171 54 Solna	
Sweden	Sweden	

Functional Testing according to D51114 Rev: D

18301

Test Description	Acceptance Criteria	Test Result
LOW Positive	HBV DETECTED: 1.26-1.70 log10 IU/mL	Passed
HIGH Positive	HBV DETECTED: 5.54-5.86 log10 IU/mL	Passed
Negative	HBV NOT DETECTED	Passed

Signature of Quality Assurance Date

Name: Samaneh Vahid

Title: QA Release Coordinator