Liofilchem® Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Mecillinam MEC 0.016-256 μg/mL	011824094	18.01.2024	2027.01.17

Ref.

92017 - 920170 - 920171

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.03-0.25	0.03

Batch Release

Approved

Date25.01.2024SignatureQuality Control

The results reported were obtained at the time of release.

(This document has been established electronically and is valid without signature)