

# CERTIFICATE OF ANALYSIS

## **POSITIVE CONTROL KIT**

PRODUCT CODE: SI 190.911	NAME: POSITIVE CONTROL KIT	
BATCH N.: PCK24003	EXPIRY DATE: 17/04/2025	
PACKING DATE: 22/02/2024	PACKED KITS No.: 30	

### **♦ RECONSTITUTING AND DILUTING SOLUTIONS**

Eugonic broth

Batch N.: FLA24001

**Expire date:** 28/01/2026

	Physical parameters *	Decontamination at 37 °C after 72 h	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%≤35%	CV%≤35%
Found	ОК	ОК	11,5%	<5%

<sup>\*</sup> Volume, transparency, colour

#### ♦ LYOPHILIZED MICROORGANISM

**Batch N.:** 23002

Expire date: 13/06/2025

	Growth inter batch (CV% ≤35%)	Physical check**
Expected	Compliant	Compliant
Found	ок	ок

<sup>\*\*</sup> Aspect

STORAGE CONDITIONS: 2 ÷ 8 °C

Date of Approval: Pividori Denis: 27/02/2024 Manufacturing Manager

Date of Release: 27/02/2024

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance

#### ALIFAX S.r.I.

Registered and head office via Petrarca, 2/1 - 35020 Polverara PD Italy Tel. (+39) 049 0992000 Fax (+39) 049 5855434 info@alifax.com - info.alifax@pec.alifax.com R&D, Design and Manufacturing Dept via Merano, 30,33045 Nimis UD Italy Tel. (+39) 0432 197900 Fax (+39) 0432 547378 info.udine@alifax.com - info.udine@pec.alifax.com

©Copyright 2022. This document is exclusive property of ALIFAX S.r.I, confidential and not freely disclosable. All rights reserved.