

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ QUALITY CONTROL CERTIFICATE	N°9137 - 9137/1 Revisione 4 del 09.10.2013 Pag. 1 di 1
-----------------------------	--	---

CARATTERISTICHE DEL PRODOTTO / PRODUCT CHARACTERISTICS

Prodotto / <i>Product</i>	AMPHOTERICIN B
Sigla / <i>Initial</i>	AMB 10 ug
Lotto / <i>Batch</i>	021121053
Data produzione / <i>Production date</i>	11.02.2021
Data scadenza / <i>Expiry date</i>	2024.02.11
Contenuto antibiotico del disco / <i>Antimicrobial Disk content</i>	Amphotericin B 10 ug

CONTROLLO QUALITÀ / QUALITY CONTROL

Terreno di coltura / <i>Culture medium</i>	Mueller Hinton+Glucose+Methylen Blue Agar
Tempo di incubazione / <i>Incubation time</i>	20-24 h in caso di crescita insufficiente prolungare l'incubazione fino a 48 ore / <i>read at 48 hours only when insufficient growth is observed after 24 hours of incubation</i>
Temperatura / <i>Temperature</i>	36 ± 1°C
Modalità di incubazione / <i>Procedure of incubation</i>	O₂
Microrganismo/ <i>Microorganism</i>	RESULTS mm
<i>C. albicans</i>	ATCC 90028 15
<i>C. krusei</i>	ATCC 6258 11
<i>C. parapsilosis</i>	ATCC 22019 15
<i>C. tropicalis</i>	ATCC 750 15

LOTTO/BATCH : ☒ Idoneo / Approved

☐ Non Idoneo / Not approved

DATA / DATE 19.02.2021

Responsabile Controllo Qualità

Quality Control Manager

(D. Vitagliano)

Dario Vitagliano