

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Amphotericin B AMB 10 µg	102622141	2025.10.25

Ref.

9137 – 9137/1 – 9137/2

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	18
<i>Candida parapsilosis</i> ATCC® 22019	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	17
<i>Candida krusei</i> ATCC® 6258	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	13
<i>Candida tropicalis</i> ATCC® 750	Mueller Hinton+ Glucose+ Methylen Blue Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	17

Batch Release

Approved

Date

04.11.2022

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.