



**ZEISS**

# EU Declaration of Conformity

We, the manufacturer, CARL ZEISS SUZHOU CO., LTD. MODERN INDUSTRIAL SQUARE 3-B, No.333, XINGPU ROAD SIP, 215126 SUZHOU, CHINA (SRN:CN-MF-000017606) , declare under our sole responsibility that the product mentioned below is in conformity with the requirements of the following regulation and directives:

- (EU) 2017/ 746 on in vitro diagnostic medical devices of April 5, 2017
- 2011/ 65/ EU of June 8, 2011 and (EU) 2015/ 863 of March 31, 2015, on the restriction of the use of hazardous substances in electrical and electronic equipment

**Authorised representative:** Carl Zeiss Microscopy GmbH, Carl Zeiss Promenade 10, 07745 Jena, Germany

Any modification to the product, not authorized by us, will invalidate this declaration.

**Product identification:**

Stereo microscope

**Trade Name:**

**Stemi 305**  
with accessories

**Stereo microscope system to visualize samples derived from the human body**

**Standards:**

EN 61010-1:	2019
EN 61010-2-101:	2017
EN 61326-1:	2013
EN 61326-2-6:	2013
EN IEC 63000 :	2018

Risk class according to Annex VIII (EU) 2017 / 746:	A
Basic-UDI-DI according to Annex VI (EU) 2017 / 746:	6909262CNA001Q9
Conformity Assessment according to:	Annex II and III (EU) 2017 / 746
RoHS-conform with exception:	6a, 6b,6b-I,6b-II,13a
Basis – Record of Conformity No.:	KC-MIK5-1041, Version 02
Registered:	CZSZ MIC CE 001-2022

The product is marked with



Date: **Suzhou, 13.07.2023**



---

**Lei Xie**  
General Manager  
Carl Zeiss Suzhou Co., Ltd.



---

**Jian Shu Wang**  
Person responsible  
acc. Article 15 (EU) 2017/ 746  
Carl Zeiss Suzhou Co.,Ltd.