

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Itraconazole ITC 8 µg	052422141	2025.05.23

Ref.

9139 – 9139/1 – 9139/2

Antimicrobial Susceptibility Testing

Tested according to in-house disc diffusion method for antifungal agents

Control strains	Medium	Inoculum	Incubation	Results Halo (mm)
<i>Candida albicans</i> ATCC® 90028	R.P.M.I Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	28
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	30
<i>Candida tropicalis</i> ATCC 750	R.P.M.I Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	30
<i>Candida krusei</i> ATCC® 6258	R.P.M.I Agar	McFarland 0.5	Aerobic , 24-48 h , 35 ± 2°C	29

Batch Release

Approved

Date 03.06.2022

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.