

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Tetracycline TE 0.016-256 µg/mL	070324173	03.07.2024	2027.07.03

Ref.

92114 – 921140 – 921141

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–1	0.12
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32	12
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	1
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32	12
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	4–32	8
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.06–0.5	0.125
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO ₂ 20-24 h	0.25–1	1
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.125–0.5	0.25
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	8
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO ₂ 16-20 h	0.25–1*	0.38

*Established and validated by EUCAST

Batch Release

Approved

Date

11.07.2024

Signature

The results reported were obtained at the time of release.

Quality Control

(This document has been
established electronically and
is valid without signature)