

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                                         | Batch     | Production date | Expiration date |
|-------------------------------------------------|-----------|-----------------|-----------------|
| Piperacillin-tazobactam TZP 0.016/4-256/4 µg/mL | 110524084 | 05.11.2024      | 2027.11.05      |

### Ref.

92108 – 921080 – 921081

### Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains                              | Medium                    | Inoculum                   | Incubation                   | Expected Results<br>MIC range (mg/L) | Results<br>MIC (mg/L) |
|----------------------------------------------|---------------------------|----------------------------|------------------------------|--------------------------------------|-----------------------|
| <i>Escherichia coli</i><br>ATCC® 25922       | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 1–8                                  | 1.5                   |
| <i>Pseudomonas aeruginosa</i><br>ATCC® 27853 | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 1–8                                  | 4                     |
| <i>Escherichia coli</i><br>ATCC® 35218       | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 0.5–2                                | 1                     |
| <i>Klebsiella pneumoniae</i><br>ATCC® 700603 | Mueller Hinton II<br>Agar | 0.5 McFarland in<br>saline | 35 ± 2°C, ambient<br>16-20 h | 8–32                                 | 8                     |

### Batch Release

Approved

Date

15.11.2024

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.