



CERTIFICATE OF ANALYSIS

MIC_COA_SI390900#SCR24152.pdf

Product Code SI 390.900	Name URO-QUICK SCREENING KIT	Quantity manufactured 27	
-----------------------------------	--	---------------------------------------	--

Lot Number SCR24152	Manufacturing Date 03/06/2024	Expiry Date 02/08/2025	CAP color Green	Printed Band White
-------------------------------	---	----------------------------------	---------------------------	------------------------------

Eugonic Broth

Lot Number **SCR24152**
Expiry Date **02/08/2025**
CAP color **Green**
Printed Band **White**

	Physical parameters (Volume, transparency, colour)	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV%<=35%	CV%<=35%
Found	OK	OK	11.5%	< 5%

Check of the presenze of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Controlled By **Cirio Alberto**
Approved By Manufacturing Manager **Pividori Denis**

Date of Control **07/06/2024**
Date of Approval **07/06/2024**

We certify that the lot in object has been manufactured and tested according with ALIFAX manufacturing and quality control procedures and specifications. We confirm that the above referenced product conformed to the specifications and then it is approved.

Approved by Quality Assurance **Zorzut Milena**

Date of Release **07/06/2024**

This is a computer-generated document. No signature is required.

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
Tel (+39) 049 0992000 - Fax (+39) 049 5855434
Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
Tel (+39) 0432 547454 - Fax (+39) 0432 547378
Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com