

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Cefiderocol FDC 30 µg	100424067	2025.10.04

Ref.

9266 – 9266/1 – 9266/2

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	25-31 (CLSI) 24-30 (EUCAST)	27
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	22-31 (CLSI) 23-29 (EUCAST)	27

Batch Release

Approved

Date

10.10.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.