

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
Neomycin N 30 µg	012125192	2028.01.21

### Ref.

9032 – 9032/1 – 9032/2

### Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results Zone range (mm)	Results Zone (mm)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	14-20*	20
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland	35 ± 1°C, ambient 16-18 h	16-22	22

\*Established and validated by EUCAST

### Batch Release

Approved

Date

28.01.2025

Signature

Quality Control  
(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.