

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product             | Batch     | Expiration date |
|---------------------|-----------|-----------------|
| Bacitracin BA 10 IU | 120624080 | 2027.12.06      |

### Ref.

9051 – 9051/1 – 9051/2

### Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains                              | Medium                    | Inoculum      | Incubation                   | Results<br>Zone (mm) |
|----------------------------------------------|---------------------------|---------------|------------------------------|----------------------|
| <i>Escherichia coli</i><br>ATCC® 25922       | Mueller Hinton II<br>Agar | 0.5 McFarland | 35 ± 1°C, ambient<br>16-18 h | 6                    |
| <i>Staphylococcus aureus</i><br>ATCC® 25923  | Mueller Hinton II<br>Agar | 0.5 McFarland | 35 ± 1°C, ambient<br>16-18 h | 20                   |
| <i>Pseudomonas aeruginosa</i><br>ATCC® 27853 | Mueller Hinton II<br>Agar | 0.5 McFarland | 35 ± 1°C, ambient<br>16-18 h | 6                    |

### Batch Release

Approved

Date

17.12.2024

Signature

Quality Control

(This document has been  
established electronically and  
is valid without signature)

The results reported were obtained at the time of release.