

• Certificate of Analysis

Product Name : MBT Sepsityper IVD Kit
Part No.: 1834338
LOT No.: 6010524013
Use by¹⁾: 2026-09-30
Date of Production: 2025-01-10
Storage Conditions : All kit components must be stored at room temperature (15–25 °C).

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Identification of <i>Candida glabrata</i> from a blood culture using the MALDI Biotyper and the MBT Sepsityper IVD Kit following the kit sample preparation procedure.	Classification Result Report lists „ <i>Candida glabrata</i> ” as the best match. log(score) Value ≥ 2.00 of the best match.	Passed Passed
Identification of blank test samples using the MALDI Biotyper and the MBT Sepsityper IVD Kit following the kit sample preparation procedure.	Classification Result Report lists „no peaks found” or “No Organism Identification Possible” as the best match. log(score) Value < 1.60 of the best match.	Passed Passed

Remarks:

¹⁾ Kit components should be handled carefully to avoid contaminations.

Bruker Daltonics GmbH & Co. KG
 Fahrenheitstr. 4
 28359 Bremen
 Germany
 Phone +49-421-2205-0

QC release date: 2025-01-15


 Quality Management